

RECEIVED

Florida Department of Law Enforcement  
Alcohol Testing Program

APR 21 2009

FDLE  
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

LD3

Agency: SARASOTA COUNTY SO  
Time of Inspection: 14:49

Date of Inspection: 04/13/2009

Serial Number: 80-000861  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 200808A Exp: 08/14/2010	0.08g/210L Test (g/210L) Lot#: 200708C Exp: 08/31/2009	0.20g/210L Test (g/210L) Lot#: 200805C Exp: 05/13/2010	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 8323011 Exp: 11/22/2010
0.000	0.048	0.079	0.197	0.080
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.079
Standard Deviations	0.0006	0.0004	0.0010	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

I Heroby Certify this to be a true and accurate duplicate of the document in the custody of the Florida Department of Law Enforcement Alcohol Testing Program

*Maria D. Brubaker*  
Custodian of Records

The above instrument complies  does not comply  with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Donald P. Suereth*  
Signature and Printed Name

I HEREBY CERTIFY THE AUTHENTICITY OF THE SEAL AFFIXED HERETO.

04/13/2009  
Date